

The Trilogy ventilator is compliant with Section 21, Emissions of RF Energy in accordance with RTCA/DO-160F testing requirements. Such certification is required for use aboard all commercial airlines landing or taking off within the United States. This certification complies with the Air Carriers Access Act and the Department of Transportation's Nondiscrimination on the Basis of Disability in Air Travel policy.

Patients traveling with Trilogy may be required by their airline to produce evidence of RTCA/DO-160F certification as well as comply with other requirements for traveling with Trilogy. Patients should check with their airlines regarding the certification and other policies before attempting to fly with the Trilogy ventilator.

RUBICOM




RTCA/DO-160F TEST REPORT
FOR THE
PHILIPS RESPIRONICS
TRILOGY, MIXED MODE VENTILATOR



SEPTEMBER 2009

Rubicom
284 West Drive, Suite B
Melbourne, Florida 32904

RTCA/DO-160F**TEST REPORT****In Accordance with Directive: RCTA/DO-160F****Manufacturer: Philips Respironics****Model: Trilogy, Mixed Mode Ventilator****Rubicom Report: JA-2533****Issue Date: September 18, 2009****Approved By:**

William Jeff Prosser
General Manager
Rubicom Systems, a division of ACS
284 West Drive, Suite B
Melbourne, Florida 32904
(321) 951-1710, Ext. 101

Customer Representative:

J. Cheng
Philips Respironics
Home Healthcare Solutions
1001 Murry Ridge Lane
Murrysville, PA 15668

Test Dates:

Start Date: 11 September 2009
Completion Date: 14 September 2009

This test report shall not be reproduced except in full. This report may be reproduced in part with prior written consent of Rubicom or ACS, Inc. The results contained in this report are representative of the sample(s) submitted for evaluation.

This report contains 26 pages